

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011519\
 Data File : FG003486.D
 Signal(s) : FID1A.ch
 Acq On : 15 Jan 2019 12:38
 Operator : AJ\SJ
 Sample : PB116368BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 15 14:30:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.474	2590844	19.968 ug/ml
Target Compounds			
1) N-OCTANE	2.118	2539562	18.833 ug/ml
2) N-DECANE	4.201	3256885	24.655 ug/ml
3) N-DODECANE	6.224	3335762	25.612 ug/ml
4) N-TETRADECANE	8.019	3352023	24.559 ug/ml
5) N-HEXADECANE	9.620	3385995	23.672 ug/ml
6) N-OCTADECANE	11.062	3438277	23.471 ug/ml
7) N-EICOSANE	12.373	3407746	23.157 ug/ml
8) N-DOCOSANE	13.573	3340448	22.881 ug/ml
10) N-TETRACOSANE	14.677	3243476	22.452 ug/ml
11) N-HEXACOSANE	15.700	3130600	21.893 ug/ml
12) N-OCTACOSANE	16.652	3043449	21.443 ug/ml
13) N-TRIACONTANE	17.540	2933679	20.822 ug/ml
14) N-DOTRIACONTANE	18.375	2731016	20.190 ug/ml
15) N-TETRATRIACONTANE	19.160	2218947	17.234 ug/ml
16) N-HEXATRIACONTANE	19.904	1615351	13.536 ug/ml
17) N-OCTATRIACONTANE	20.621	1257719	11.723 ug/ml
18) N-TETRACONTANE	21.473	1129528	10.682 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011519\
Data File : FG003486.D
Signal(s) : FID1A.ch
Acq On : 15 Jan 2019 12:38
Operator : AJ\SJ
Sample : PB116368BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 15 14:30:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

