

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012119\
 Data File : FG003548.D
 Signal(s) : FID1A.ch
 Acq On : 21 Jan 2019 10:37
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 21 23:53:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.484	6017841	46.380 ug/ml
Target Compounds			
2) N-DECANE	4.200	7393429	55.969 ug/ml
3) N-DODECANE	6.226	7372516	56.605 ug/ml
4) N-TETRADECANE	8.023	7422775	54.384 ug/ml
5) N-HEXADECANE	9.626	7440336	52.016 ug/ml
6) N-OCTADECANE	11.069	7397245	50.496 ug/ml
7) N-EICOSANE	12.381	7206192	48.969 ug/ml
8) N-DOCOSANE	13.582	6968053	47.729 ug/ml
10) N-TETRACOSANE	14.688	6711572	46.459 ug/ml
11) N-HEXACOSANE	15.711	6490079	45.387 ug/ml
12) N-OCTACOSANE	16.662	6289810	44.316 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012119\
Data File : FG003548.D
Signal(s) : FID1A.ch
Acq On : 21 Jan 2019 10:37
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 21 23:53:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

