

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012221\
Data File : FG007573.D
Signal(s) : FID1A.ch
Acq On : 22 Jan 2021 19:57
Operator : DD\AJ
Sample : M1219-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 22 21:46:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010421.M
Quant Title :
QLast Update : Tue Jan 05 02:18:25 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.634	2709838	22.612 ug/ml
Target Compounds				
10)	N-TETRACOSANE	14.883	215162	1.573 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012221\
Data File : FG007573.D
Signal(s) : FID1A.ch
Acq On : 22 Jan 2021 19:57
Operator : DD\AJ
Sample : M1219-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 22 21:46:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010421.M
Quant Title :
QLast Update : Tue Jan 05 02:18:25 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.634 min
Delta R.T.: -0.014 min
Response: 2709838
Conc: 22.61 ug/ml

Instrument :
FID_G
ClientSampleId :



#10 N-TETRACOSANE

R.T.: 14.883 min
Delta R.T.: 0.027 min
Response: 215162
Conc: 1.57 ug/ml