

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

Report

Area Percent
Manual IntegrationsAPPROVED
Reviewed By :Yogesh Patel 01/24/2025
Supervised By :Ankita Jodhani 01/27/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Da
Data File : FG015146.D
Signal(s) : FID1A.ch
Acq On : 23 Jan 2025 21:22
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG011325.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.007	1.973	2.072	BB	551238	5921969	81.71%	5.400%
2	4.543	4.507	4.583	BB	627703	6404343	88.36%	5.839%
3	6.726	6.684	6.766	BB	679509	6779644	93.54%	6.182%
4	8.562	8.527	8.599	BB	646548	6759321	93.26%	6.163%
5	10.178	10.136	10.218	BB	641701	6978306	96.28%	6.363%
6	11.627	11.589	11.665	BB	651260	7247651	100.00%	6.608%
7	12.943	12.903	12.988	BB	587706	7096578	97.92%	6.471%
8	14.146	14.096	14.189	BB	573007	6950104	95.89%	6.337%
9	15.049	15.004	15.078	BV	459589	6084217	83.95%	5.547%
10	15.254	15.215	15.300	BB	534347	6874089	94.85%	6.268%
11	16.278	16.236	16.321	BB	488209	6707820	92.55%	6.116%
12	17.232	17.180	17.278	BB	499222	6554555	90.44%	5.976%
13	18.121	18.070	18.163	BB	454886	6471698	89.29%	5.901%
14	18.957	18.899	19.003	BB	439998	6193721	85.46%	5.647%
15	19.741	19.685	19.778	BB	371104	5315751	73.34%	4.847%
16	20.481	20.430	20.526	BB	290153	4262813	58.82%	3.887%
17	21.230	21.175	21.293	BB	208255	3675571	50.71%	3.351%
18	22.170	22.095	22.242	BB	138273	3396851	46.87%	3.097%
Sum of corrected areas:						109675000		

FG011325.M Fri Jan 24 04:55:05 2025