

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012419\
 Data File : FG003585.D
 Signal(s) : FID1A.ch
 Acq On : 24 Jan 2019 18:19
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 25 03:47:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.483	6702512	51.657 ug/ml
Target Compounds			
2) N-DECANE	4.203	8444543	63.926 ug/ml
3) N-DODECANE	6.229	8379380	64.336 ug/ml
4) N-TETRADECANE	8.025	8314399	60.916 ug/ml
5) N-HEXADECANE	9.627	8264950	57.781 ug/ml
6) N-OCTADECANE	11.070	8153912	55.662 ug/ml
7) N-EICOSANE	12.382	7913564	53.775 ug/ml
8) N-DOCOSANE	13.582	7657604	52.452 ug/ml
10) N-TETRACOSANE	14.688	7421626	51.374 ug/ml
11) N-HEXACOSANE	15.709	7303194	51.073 ug/ml
12) N-OCTACOSANE	16.661	7239484	51.007 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012419\
Data File : FG003585.D
Signal(s) : FID1A.ch
Acq On : 24 Jan 2019 18:19
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 25 03:47:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

