

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012519\
 Data File : FG003588.D
 Signal(s) : FID1A.ch
 Acq On : 25 Jan 2019 09:10
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 25 12:40:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.482	6409008	49.395 ug/ml
Target Compounds			
1) N-OCTANE	2.113	7560701	56.069 ug/ml
2) N-DECANE	4.203	8095207	61.281 ug/ml
3) N-DODECANE	6.228	8042791	61.752 ug/ml
4) N-TETRADECANE	8.024	7987151	58.519 ug/ml
5) N-HEXADECANE	9.626	7926582	55.416 ug/ml
6) N-OCTADECANE	11.069	7820833	53.388 ug/ml
7) N-EICOSANE	12.381	7588166	51.564 ug/ml
8) N-DOCOSANE	13.580	7342978	50.297 ug/ml
10) N-TETRACOSANE	14.686	7125147	49.322 ug/ml
11) N-HEXACOSANE	15.710	6992817	48.903 ug/ml
12) N-OCTACOSANE	16.661	6935320	48.864 ug/ml
13) N-TRIACONTANE	17.548	6931937	49.200 ug/ml
14) N-DOTRIACONTANE	18.385	6731146	49.763 ug/ml
15) N-TETRATRIACONTANE	19.171	6574727	51.065 ug/ml
16) N-HEXATRIACONTANE	19.913	6258077	52.440 ug/ml
17) N-OCTATRIACONTANE	20.631	5662205	52.777 ug/ml
18) N-TETRACONTANE	21.489	5616615	53.115 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012519\
Data File : FG003588.D
Signal(s) : FID1A.ch
Acq On : 25 Jan 2019 09:10
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 25 12:40:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

