

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012519\
 Data File : FG003593.D
 Signal(s) : FID1A.ch
 Acq On : 25 Jan 2019 12:19
 Operator : AJ\SJ
 Sample : K1216-01MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 28 10:02:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.504	2173733	16.753 ug/mlm
Target Compounds			
1) N-OCTANE	2.111	1079034	8.002 ug/ml
2) N-DECANE	4.203	2416676	18.294 ug/ml
3) N-DODECANE	6.228	2264580	17.387 ug/ml
4) N-TETRADECANE	8.023	3357028	24.596 ug/mlm
5) N-HEXADECANE	9.626	5480488	38.315 ug/ml
6) N-OCTADECANE	11.071	4137730	28.246 ug/ml
7) N-EICOSANE	12.389	4120385	27.999 ug/ml
8) N-DOCOSANE	13.593	2676975	18.336 ug/mlm
10) N-TETRACOSANE	14.670	1808144	12.516 ug/mlm
11) N-HEXACOSANE	15.728	3417672	23.901 ug/mlm
12) N-OCTACOSANE	16.680	4719443	33.252 ug/mlm
13) N-TRIACONTANE	17.573	6094336	43.255 ug/mlm
14) N-DOTRIACONTANE	18.386	4120867	30.465 ug/mlm
15) N-TETRATRIACONTANE	19.149	315512	2.451 ug/ml
16) N-HEXATRIACONTANE	19.895	1123832	9.417 ug/ml
17) N-OCTATRIACONTANE	20.663	2577294	24.023 ug/ml
18) N-TETRACONTANE	21.523	2879585	27.231 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012519\
Data File : FG003593.D
Signal(s) : FID1A.ch
Acq On : 25 Jan 2019 12:19
Operator : AJ\SJ
Sample : K1216-01MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 28 10:02:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

