

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012621\
 Data File : FG007584.D
 Signal(s) : FID1A.ch
 Acq On : 26 Jan 2021 17:05
 Operator : DD\AJ
 Sample : M1234-01MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 HOLLAND-TUNNEL-TEST-PIT-5MS

Manual Integrations
 APPROVED

mohammad
 1/27/2021 5:23:32 PM

Integration File: Sample.e
 Quant Time: Jan 27 00:30:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010421.M
 Quant Title :
 QLast Update : Tue Jan 05 02:18:25 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.639	3496125	29.173 ug/mlm
Target Compounds			
1) N-OCTANE	2.241	2269487	15.478 ug/ml
2) N-DECANE	4.338	2890498	19.212 ug/ml
3) N-DODECANE	6.363	3039914	20.043 ug/ml
4) N-TETRADECANE	8.163	3367646	22.202 ug/ml
5) N-HEXADECANE	9.769	3277341	21.934 ug/ml
6) N-OCTADECANE	11.216	3156804	21.165 ug/ml
7) N-EICOSANE	12.533	4152523	28.491 ug/ml
8) N-DOCOSANE	13.735	5453029	38.947 ug/mlm
10) N-TETRACOSANE	14.827	13171239	96.300 ug/ml
11) N-HEXACOSANE	15.871	3907750	29.754 ug/ml
12) N-OCTACOSANE	16.826	6600514	51.970 ug/ml
13) N-TRIACONTANE	17.715	3474543	28.244 ug/mlm
14) N-DOTRIACONTANE	18.549	3027192	26.721 ug/mlm
15) N-TETRATRIACONTANE	19.337	2970761	31.516 ug/mlm
16) N-HEXATRIACONTANE	20.079	2710474	36.959 ug/mlm
17) N-OCTATRIACONTANE	20.815	2203084	34.589 ug/mlm
18) N-TETRACONTANE	21.729	2282773	38.721 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012621\
Data File : FG007584.D
Signal(s) : FID1A.ch
Acq On : 26 Jan 2021 17:05
Operator : DD\AJ
Sample : M1234-01MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
HOLLAND-TUNNEL-TEST-PIT-5MS

Manual Integrations
APPROVED

mohammad
1/27/2021 5:23:32 PM

Integration File: Sample.e
Quant Time: Jan 27 00:30:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010421.M
Quant Title :
QLast Update : Tue Jan 05 02:18:25 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

