

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012623\
 Data File : FG011311.D
 Signal(s) : FID1A.ch
 Acq On : 26 Jan 2023 12:23
 Operator : YP\AJ
 Sample : 01243-04MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 26 14:08:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG123022.M
 Quant Title :
 QLast Update : Mon Jan 02 04:44:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.083	1986250	12.961 ug/ml
Target Compounds			
1) N-OCTANE	2.073	1919271	11.630 ug/ml
2) N-DECANE	4.594	2331758	13.454 ug/ml
3) N-DODECANE	6.766	2418767	13.710 ug/ml
4) N-TETRADECANE	8.599	2444508	13.709 ug/ml
5) N-HEXADECANE	10.214	2495960	13.760 ug/ml
6) N-OCTADECANE	11.662	2490139	13.565 ug/ml
7) N-EICOSANE	12.978	2455525	13.672 ug/ml
8) N-DOCOSANE	14.180	2372166	13.422 ug/ml
10) N-TETRACOSANE	15.288	2364049	13.548 ug/ml
11) N-HEXACOSANE	16.314	2310442	13.563 ug/ml
12) N-OCTACOSANE	17.267	2194884	13.346 ug/ml
13) N-TRIACONTANE	18.156	2154537	13.488 ug/ml
14) N-DOTRIACONTANE	18.993	2057788	13.429 ug/ml
15) N-TETRATRIACONTANE	19.779	1816689	12.708 ug/ml
16) N-HEXATRIACONTANE	20.521	1638687	12.344 ug/ml
17) N-OCTATRIACONTANE	21.280	1640840	12.381 ug/ml
18) N-TETRACONTANE	22.232	1698584	12.678 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012623\
Data File : FG011311.D
Signal(s) : FID1A.ch
Acq On : 26 Jan 2023 12:23
Operator : YP\AJ
Sample : 01243-04MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 26 14:08:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG123022.M
Quant Title :
QLast Update : Mon Jan 02 04:44:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

