

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012820\
 Data File : FG006559.D
 Signal(s) : FID1A.ch
 Acq On : 28 Jan 2020 18:39
 Operator : DD\AJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 29 02:44:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
 Quant Title :
 QLast Update : Wed Jan 29 02:43:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.583	8999243	100.911 ug/ml
Target Compounds			
1) N-OCTANE	2.555	9035113	104.390 ug/ml
2) N-DECANE	4.611	9416117	102.981 ug/ml
3) N-DODECANE	6.572	9826950	102.067 ug/ml
4) N-TETRADECANE	8.311	10104826	101.218 ug/ml
5) N-HEXADECANE	9.864	10350269	100.444 ug/ml
6) N-OCTADECANE	11.264	10328480	99.977 ug/ml
7) N-EICOSANE	12.540	10417492	99.821 ug/ml
8) N-DOCOSANE	13.708	10361214	100.304 ug/ml
10) N-TETRACOSANE	14.784	10355789	100.953 ug/ml
11) N-HEXACOSANE	15.780	10326908	101.313 ug/ml
12) N-OCTACOSANE	16.707	10367160	100.608 ug/ml
13) N-TRIACONTANE	17.572	10332428	99.071 ug/ml
14) N-DOTRIACONTANE	18.384	9896218	99.739 ug/ml
15) N-TETRATRIACONTANE	19.147	9685244	99.336 ug/ml
16) N-HEXATRIACONTANE	19.868	8998801	98.592 ug/ml
17) N-OCTATRIACONTANE	20.559	8428403	98.964 ug/ml
18) N-TETRACONTANE	21.359	8293246	101.402 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012820\
Data File : FG006559.D
Signal(s) : FID1A.ch
Acq On : 28 Jan 2020 18:39
Operator : DD\AJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 29 02:44:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
Quant Title :
QLast Update : Wed Jan 29 02:43:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

