

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012820\
 Data File : FG006562.D
 Signal(s) : FID1A.ch
 Acq On : 28 Jan 2020 20:09
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 29 02:45:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
 Quant Title :
 QLast Update : Wed Jan 29 02:43:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.578	948605	10.637 ug/ml
Target Compounds			
1) N-OCTANE	2.557	868673	10.036 ug/ml
2) N-DECANE	4.609	905924	9.908 ug/ml
3) N-DODECANE	6.567	945513	9.820 ug/ml
4) N-TETRADECANE	8.306	991828	9.935 ug/ml
5) N-HEXADECANE	9.858	1052105	10.210 ug/ml
6) N-OCTADECANE	11.258	1072306	10.380 ug/ml
7) N-EICOSANE	12.532	1099901	10.539 ug/ml
8) N-DOCOSANE	13.699	1095132	10.602 ug/ml
10) N-TETRACOSANE	14.776	1089728	10.623 ug/ml
11) N-HEXACOSANE	15.772	1087053	10.665 ug/ml
12) N-OCTACOSANE	16.698	1143660	11.099 ug/ml
13) N-TRIACONTANE	17.563	1235249	11.844 ug/ml
14) N-DOTRIACONTANE	18.374	1111364	11.201 ug/ml
15) N-TETRATRIACONTANE	19.138	1046971	10.738 ug/ml
16) N-HEXATRIACONTANE	19.859	977531	10.710 ug/ml
17) N-OCTATRIACONTANE	20.551	908578	10.668 ug/ml
18) N-TETRACONTANE	21.351	890876	10.893 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012820\
Data File : FG006562.D
Signal(s) : FID1A.ch
Acq On : 28 Jan 2020 20:09
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 29 02:45:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
Quant Title :
QLast Update : Wed Jan 29 02:43:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

