

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012820\
 Data File : FG006563.D
 Signal(s) : FID1A.ch
 Acq On : 28 Jan 2020 20:39
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 29 02:45:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
 Quant Title :
 QLast Update : Wed Jan 29 02:43:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.576	504060	5.652 ug/ml
Target Compounds			
1) N-OCTANE	2.558	471653	5.449 ug/ml
2) N-DECANE	4.609	487047	5.327 ug/ml
3) N-DODECANE	6.567	508546	5.282 ug/ml
4) N-TETRADECANE	8.306	530637	5.315 ug/ml
5) N-HEXADECANE	9.857	561819	5.452 ug/ml
6) N-OCTADECANE	11.258	575070	5.567 ug/ml
7) N-EICOSANE	12.532	591839	5.671 ug/ml
8) N-DOCOSANE	13.699	587306	5.686 ug/ml
10) N-TETRACOSANE	14.775	583330	5.687 ug/ml
11) N-HEXACOSANE	15.772	586891	5.758 ug/ml
12) N-OCTACOSANE	16.697	640930	6.220 ug/mlm
13) N-TRIACONTANE	17.562	738433	7.080 ug/mlm
14) N-DOTRIACONTANE	18.374	649394	6.545 ug/mlm
15) N-TETRATRIACONTANE	19.138	579986	5.949 ug/ml
16) N-HEXATRIACONTANE	19.859	563061	6.169 ug/ml
17) N-OCTATRIACONTANE	20.550	500659	5.879 ug/ml
18) N-TETRACONTANE	21.349	505754	6.184 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012820\
Data File : FG006563.D
Signal(s) : FID1A.ch
Acq On : 28 Jan 2020 20:39
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 29 02:45:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
Quant Title :
QLast Update : Wed Jan 29 02:43:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

