

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020123\
 Data File : FG011347.D
 Signal(s) : FID1A.ch
 Acq On : 01 Feb 2023 11:49
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 02 05:17:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG013023.M
 Quant Title :
 QLast Update : Mon Jan 30 21:53:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.077	5856265	44.370 ug/ml
Target Compounds			
1) N-OCTANE	2.076	5896491	44.783 ug/ml
2) N-DECANE	4.595	6253892	45.004 ug/ml
3) N-DODECANE	6.766	6411310	45.102 ug/ml
4) N-TETRADECANE	8.598	6486056	45.231 ug/ml
5) N-HEXADECANE	10.211	6680416	45.251 ug/ml
6) N-OCTADECANE	11.658	6753060	45.124 ug/ml
7) N-EICOSANE	12.972	6765377	44.924 ug/ml
8) N-DOCOSANE	14.175	6574321	44.739 ug/ml
10) N-TETRACOSANE	15.282	6569779	44.498 ug/ml
11) N-HEXACOSANE	16.306	6422693	44.334 ug/ml
12) N-OCTACOSANE	17.259	6237402	43.885 ug/ml
13) N-TRIACONTANE	18.149	5977496	42.876 ug/ml
14) N-DOTRIACONTANE	18.983	5548069	39.913 ug/ml
15) N-TETRATRIACONTANE	19.768	5028262	40.364 ug/ml
16) N-HEXATRIACONTANE	20.511	4784300	41.236 ug/ml
17) N-OCTATRIACONTANE	21.268	5145287	46.581 ug/ml
18) N-TETRACONTANE	22.215	5233027	46.530 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020123\
Data File : FG011347.D
Signal(s) : FID1A.ch
Acq On : 01 Feb 2023 11:49
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 02 05:17:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG013023.M
Quant Title :
QLast Update : Mon Jan 30 21:53:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

