

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020124\
 Data File : FG013399.D
 Signal(s) : FID1A.ch
 Acq On : 01 Feb 2024 10:57
 Operator : YP\AJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 01 13:57:02 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020124.M
 Quant Title :
 QLast Update : Thu Feb 01 13:56:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.078	11120788	98.700 ug/ml
Target Compounds			
1) N-OCTANE	2.058	11600724	97.805 ug/ml
2) N-DECANE	4.588	11665649	98.017 ug/ml
3) N-DODECANE	6.761	12110824	98.138 ug/ml
4) N-TETRADECANE	8.593	11917240	98.453 ug/ml
5) N-HEXADECANE	10.208	12216921	98.294 ug/ml
6) N-OCTADECANE	11.656	12696032	98.414 ug/ml
7) N-EICOSANE	12.972	12464776	98.508 ug/ml
8) N-DOCOSANE	14.175	12284382	98.562 ug/ml
10) N-TETRACOSANE	15.283	12130744	98.653 ug/ml
11) N-HEXACOSANE	16.307	11801878	98.748 ug/ml
12) N-OCTACOSANE	17.261	11449012	98.763 ug/ml
13) N-TRIACONTANE	18.151	11051571	98.725 ug/ml
14) N-DOTRIACONTANE	18.987	10294306	98.552 ug/ml
15) N-TETRATRIACONTANE	19.773	8208366	98.466 ug/ml
16) N-HEXATRIACONTANE	20.514	6194807	98.616 ug/ml
17) N-OCTATRIACONTANE	21.270	5400600	98.405 ug/ml
18) N-TETRACONTANE	22.216	5135742	97.622 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020124\
Data File : FG013399.D
Signal(s) : FID1A.ch
Acq On : 01 Feb 2024 10:57
Operator : YP\AJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 01 13:57:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020124.M
Quant Title :
QLast Update : Thu Feb 01 13:56:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

