

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020124\
 Data File : FG013404.D
 Signal(s) : FID1A.ch
 Acq On : 01 Feb 2024 13:52
 Operator : YP\AJ
 Sample : FG020124ICV
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 01 14:08:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020124.M
 Quant Title :
 QLast Update : Thu Feb 01 14:07:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.073	5703205	50.941 ug/ml
Target Compounds			
1) N-OCTANE	2.058	6022397	50.850 ug/ml
2) N-DECANE	4.584	6041150	50.783 ug/ml
3) N-DODECANE	6.756	6256178	50.747 ug/ml
4) N-TETRADECANE	8.590	6132354	50.751 ug/ml
5) N-HEXADECANE	10.203	6288235	50.832 ug/ml
6) N-OCTADECANE	11.651	6525076	50.856 ug/ml
7) N-EICOSANE	12.967	6398904	50.883 ug/ml
8) N-DOCOSANE	14.171	6298795	50.899 ug/ml
10) N-TETRACOSANE	15.279	6218619	50.884 ug/ml
11) N-HEXACOSANE	16.303	6048881	50.922 ug/ml
12) N-OCTACOSANE	17.256	5872921	51.023 ug/ml
13) N-TRIACONTANE	18.147	5671378	50.811 ug/ml
14) N-DOTRIACONTANE	18.983	5250989	50.262 ug/ml
15) N-TETRATRIACONTANE	19.768	4084800	49.030 ug/ml
16) N-HEXATRIACONTANE	20.510	2971318	47.173 ug/ml
17) N-OCTATRIACONTANE	21.266	2482812	45.273 ug/ml
18) N-TETRACONTANE	22.213	2299628	43.905 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020124\
Data File : FG013404.D
Signal(s) : FID1A.ch
Acq On : 01 Feb 2024 13:52
Operator : YP\AJ
Sample : FG020124ICV
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 01 14:08:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020124.M
Quant Title :
QLast Update : Thu Feb 01 14:07:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

