

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG013438.D
Data File : FG013438.D
Signal(s) : FID1A.ch
Acq On : 02 Feb 2024 20:12
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Reviewed By :Ankita Jodhani 02/05/2024
Supervised By :Sohil Jodhani 02/05/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020124.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.059	2.018	2.126	BB	527245	6132434	90.80%	6.007%
2	4.586	4.547	4.631	BB	612813	6273739	92.89%	6.146%
3	6.758	6.721	6.795	BB	639071	6524544	96.60%	6.391%
4	8.591	8.555	8.633	BB	607235	6386216	94.55%	6.256%
5	10.204	10.158	10.253	BB	600447	6534509	96.75%	6.401%
6	11.653	11.614	11.689	BB	583308	6754007	100.00%	6.616%
7	12.968	12.918	13.013	BB	562306	6596911	97.67%	6.462%
8	14.171	14.134	14.213	VB	527602	6480456	95.95%	6.348%
9	15.074	15.016	15.106	BV	410950	5840092	86.47%	5.721%
10	15.279	15.210	15.329	BB	488060	6396867	94.71%	6.266%
11	16.304	16.252	16.348	BB	472830	6217693	92.06%	6.091%
12	17.258	17.176	17.317	BB	455727	6074024	89.93%	5.950%
13	18.149	18.082	18.188	BV	420870	5913541	87.56%	5.793%
14	18.985	18.928	19.031	M	402250	5575254	82.55%	5.461%
15	19.770	19.724	19.807	PV	314163	4536886	67.17%	4.444%
16	20.511	20.472	20.549	VV	234959	3657518	54.15%	3.583%
17	21.266	21.225	21.342	VV	169245	3042899	45.05%	2.981%
18	22.212	22.153	22.285	M	121792	3145676	46.57%	3.081%
Sum of corrected areas:						102083265		

FG020124.M Mon Feb 05 05:21:38 2024