

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020519\
 Data File : FG003668.D
 Signal(s) : FID1A.ch
 Acq On : 05 Feb 2019 16:48
 Operator : AJ\SJ
 Sample : FG020519ICV
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 06 08:44:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
 Quant Title :
 QLast Update : Wed Feb 06 08:41:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.473	6388041	51.439 ug/ml
Target Compounds			
1) N-OCTANE	2.112	7395421	50.471 ug/ml
2) N-DECANE	4.196	7521299	50.577 ug/ml
3) N-DODECANE	6.221	7527031	50.633 ug/ml
4) N-TETRADECANE	8.016	7527131	50.808 ug/ml
5) N-HEXADECANE	9.618	7548723	50.949 ug/ml
6) N-OCTADECANE	11.061	7565441	51.136 ug/ml
7) N-EICOSANE	12.372	7447670	51.207 ug/ml
8) N-DOCOSANE	13.573	7284995	51.231 ug/ml
10) N-TETRACOSANE	14.676	7107535	51.388 ug/ml
11) N-HEXACOSANE	15.699	6939190	51.494 ug/ml
12) N-OCTACOSANE	16.650	6763082	51.592 ug/ml
13) N-TRIACONTANE	17.539	6570388	51.812 ug/ml
14) N-DOTRIACONTANE	18.374	6088061	52.147 ug/ml
15) N-TETRATRIACONTANE	19.160	5615283	52.060 ug/ml
16) N-HEXATRIACONTANE	19.901	5133799	51.519 ug/ml
17) N-OCTATRIACONTANE	20.618	4653170	50.244 ug/ml
18) N-TETRACONTANE	21.470	4733066	50.737 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020519\
Data File : FG003668.D
Signal(s) : FID1A.ch
Acq On : 05 Feb 2019 16:48
Operator : AJ\SJ
Sample : FG020519ICV
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 06 08:44:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
Quant Title :
QLast Update : Wed Feb 06 08:41:02 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

