

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020620\
 Data File : FG006614.D
 Signal(s) : FID1A.ch
 Acq On : 06 Feb 2020 11:04
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 07 07:41:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
 Quant Title :
 QLast Update : Wed Jan 29 08:10:21 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.605	3963792	42.271 ug/ml
Target Compounds			
1) N-OCTANE	2.594	4298251	48.159 ug/ml
2) N-DECANE	4.650	4546650	48.747 ug/ml
3) N-DODECANE	6.607	4698700	48.142 ug/ml
4) N-TETRADECANE	8.344	4715828	46.433 ug/ml
5) N-HEXADECANE	9.893	4754215	44.791 ug/ml
6) N-OCTADECANE	11.291	4678720	43.574 ug/ml
7) N-EICOSANE	12.563	4671163	42.703 ug/ml
8) N-DOCOSANE	13.730	4591831	42.295 ug/ml
10) N-TETRACOSANE	14.805	4543722	42.078 ug/ml
11) N-HEXACOSANE	15.800	4532304	42.082 ug/ml
12) N-OCTACOSANE	16.724	4559970	40.766 ug/ml
13) N-TRIACONTANE	17.588	4579029	38.586 ug/ml
14) N-DOTRIACONTANE	18.397	4318753	39.532 ug/ml
15) N-TETRATRIACONTANE	19.159	3959280	38.204 ug/ml
16) N-HEXATRIACONTANE	19.879	3635821	37.146 ug/ml
17) N-OCTATRIACONTANE	20.570	3432560	38.090 ug/ml
18) N-TETRACONTANE	21.371	3509722	39.502 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG020620\
Data File : FG006614.D
Signal(s) : FID1A.ch
Acq On : 06 Feb 2020 11:04
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 07 07:41:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG012820.M
Quant Title :
QLast Update : Wed Jan 29 08:10:21 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

