

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG021219\
 Data File : FG003768.D
 Signal(s) : FID1A.ch
 Acq On : 12 Feb 2019 17:34
 Operator : AJ\SJ
 Sample : K1459-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Feb 13 00:18:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
 Quant Title :
 QLast Update : Wed Feb 06 08:50:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.460	1550726	11.465 ug/ml
Target Compounds			
4) N-TETRADECANE	8.063	339570	2.151 ug/ml
12) N-OCTACOSANE	16.635	257984	1.791 ug/ml
13) N-TRIACONTANE	17.525	360693	2.549 ug/ml
14) N-DOTRIACONTANE	18.355	277889	2.122 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG021219\
Data File : FG003768.D
Signal(s) : FID1A.ch
Acq On : 12 Feb 2019 17:34
Operator : AJ\SJ
Sample : K1459-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Feb 13 00:18:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
Quant Title :
QLast Update : Wed Feb 06 08:50:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#4 N-TETRADECANE


R.T.: 8.063 min
Delta R.T.: 0.047 min
Response: 339570
Conc: 2.15 ug/ml

Instrument :
FID_G
ClientSampleId :




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.460 min
Delta R.T.: -0.013 min
Response: 1550726
Conc: 11.47 ug/ml




#12 N-OCTACOSANE

R.T.: 16.635 min
Delta R.T.: -0.017 min
Response: 257984
Conc: 1.79 ug/ml



#13 N-TRIACONTANE

R.T.: 17.525 min
Delta R.T.: -0.016 min
Response: 360693
Conc: 2.55 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.355 min
Delta R.T.: -0.020 min
Response: 277889
Conc: 2.12 ug/ml

Instrument :
FID_G
ClientSampleId :