

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022119\
 Data File : FG003923.D
 Signal(s) : FID1A.ch
 Acq On : 21 Feb 2019 22:19
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 22 00:24:28 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
 Quant Title :
 QLast Update : Wed Feb 06 08:50:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.458	5953036	44.013 ug/ml
Target Compounds			
2) N-DECANE	4.185	8216153	52.092 ug/ml
3) N-DODECANE	6.209	8032754	50.876 ug/mlm
4) N-TETRADECANE	8.004	7809423	49.479 ug/ml
5) N-HEXADECANE	9.605	7570707	48.058 ug/ml
6) N-OCTADECANE	11.048	7338207	46.616 ug/ml
7) N-EICOSANE	12.358	7048838	45.007 ug/ml
8) N-DOCOSANE	13.558	6788292	43.902 ug/ml
10) N-TETRACOSANE	14.663	6607956	43.856 ug/ml
11) N-HEXACOSANE	15.685	6500526	44.192 ug/ml
12) N-OCTACOSANE	16.635	6538738	45.385 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022119\
Data File : FG003923.D
Signal(s) : FID1A.ch
Acq On : 21 Feb 2019 22:19
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 22 00:24:28 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
Quant Title :
QLast Update : Wed Feb 06 08:50:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

