

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022223\
Data File : FG011432.D
Signal(s) : FID1A.ch
Acq On : 22 Feb 2023 16:23
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 23 01:47:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG022123.M
Quant Title :
QLast Update : Tue Feb 21 16:44:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.066	2002138	19.035 ug/ml
Target Compounds			
16) N-HEXATRIACONTANE	20.542	67391	0.753 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022223\
Data File : FG011432.D
Signal(s) : FID1A.ch
Acq On : 22 Feb 2023 16:23
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 23 01:47:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG022123.M
Quant Title :
QLast Update : Tue Feb 21 16:44:49 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um





#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 15.066 min
Delta R.T.: -0.010 min
Response: 2002138
Conc: 19.04 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.542 min
Delta R.T.: 0.031 min
Response: 67391
Conc: 0.75 ug/ml