

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022520\
 Data File : FG006732.D
 Signal(s) : FID1A.ch
 Acq On : 25 Feb 2020 13:22
 Operator : DD\AJ
 Sample : 8800 TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 25 16:30:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\C44-FG022520.M
 Quant Title :
 QLast Update : Tue Feb 25 16:27:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.661	8113289	94.742 ug/ml
Target Compounds			
1) N-OCTANE	2.706	41863292	819.587 ug/ml
2) N-NONANE	3.722	43451836	823.395 ug/ml
3) N-DECANE	4.773	67733192	824.885 ug/ml
4) N-UNDECANE	5.777	69640699	1236.574 ug/ml
5) N-DODECANE	6.723	70904998	1230.150 ug/ml
6) N-TETRADECANE	8.453	74465238	1210.242 ug/ml
7) N-HEXADECANE	9.993	63933058	991.843 ug/ml
8) N-OCTADECANE	11.373	32378557	489.410 ug/ml
9) N-EICOSANE	12.630	13038211	193.623 ug/ml
10) N-TETRACOSANE	14.863	13033465	189.386 ug/ml
11) N-OCTACOSANE	16.771	6733884	91.104 ug/ml
12) N-DOTRIACONTANE	18.437	5998947	91.907 ug/ml
14) N-HEXATRIACONTANE	19.911	4976737	95.634 ug/ml
15) N-TETRACONTANE	21.408	4470485	98.762 ug/ml
16) N-TETRTETRACOOCTANE	23.798	3564602	91.110 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022520\
Data File : FG006732.D
Signal(s) : FID1A.ch
Acq On : 25 Feb 2020 13:22
Operator : DD\AJ
Sample : 8800 TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 25 16:30:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\C44-FG022520.M
Quant Title :
QLast Update : Tue Feb 25 16:27:47 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

