

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_G\Data\FG022619\  
Data File : FG003952.D  
Signal(s) : FID1A.ch  
Acq On : 26 Feb 2019 10:51  
Operator : AJ\SJ  
Sample : RT MARKER  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Feb 27 00:32:55 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_G\Method\FG020519.M  
Quant Title :  
QLast Update : Wed Feb 06 08:50:01 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um

| Compound                    | R.T.   | Response | Conc Units   |
|-----------------------------|--------|----------|--------------|
| -----                       |        |          |              |
| System Monitoring Compounds |        |          |              |
| Target Compounds            |        |          |              |
| 3) N-DODECANE               | 6.207  | 2516021  | 15.935 ug/ml |
| 12) N-OCTACOSANE            | 16.634 | 2285702  | 15.865 ug/ml |
| -----                       |        |          |              |


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_G\Data\FG022619\  
Data File : FG003952.D  
Signal(s) : FID1A.ch  
Acq On : 26 Feb 2019 10:51  
Operator : AJ\SJ  
Sample : RT MARKER  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Feb 27 00:32:55 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_G\Method\FG020519.M  
Quant Title :  
QLast Update : Wed Feb 06 08:50:01 2019  
Response via : Initial Calibration  
Integrator: ChemStation


Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.207 min  
Delta R.T.: -0.015 min  
Response: 2516021  
Conc: 15.94 ug/ml



#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min  
Exp R.T.: 14.472 min  
Response: 0  
Conc: N.D.



#12 N-OCTACOSANE

R.T.: 16.634 min  
Delta R.T.: -0.017 min  
Response: 2285702  
Conc: 15.86 ug/ml