

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022620\
 Data File : FG006751.D
 Signal(s) : FID1A.ch
 Acq On : 26 Feb 2020 16:21
 Operator : DD\AJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 27 01:22:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\C44-FG022520.M
 Quant Title :
 QLast Update : Tue Feb 25 16:35:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
13) S	TETRACOSANE-d50 (SURR...	14.661	4857844	59.518 ug/ml
Target Compounds				
1)	N-OCTANE	2.712	22588678	445.506 ug/ml
2)	N-NONANE	3.724	23969050	460.222 ug/ml
3)	N-DECANE	4.771	37899218	469.739 ug/ml
4)	N-UNDECANE	5.774	39403659	717.369 ug/ml
5)	N-DODECANE	6.720	40520232	725.664 ug/ml
6)	N-TETRADECANE	8.448	43130771	727.599 ug/ml
7)	N-HEXADECANE	9.988	37263091	597.841 ug/ml
8)	N-OCTADECANE	11.370	18948438	296.275 ug/ml
9)	N-EICOSANE	12.630	7654000	118.226 ug/ml
10)	N-TETRACOSANE	14.862	7822641	119.108 ug/ml
11)	N-OCTACOSANE	16.770	3947597	49.469 ug/ml
12)	N-DOTRIACONTANE	18.438	3588288	51.621 ug/ml
14)	N-HEXATRIACONTANE	19.912	2792700	52.987 ug/ml
15)	N-TETRACONTANE	21.412	2413334	52.220 ug/ml
16)	N-TETRTETRACOOCTANE	23.807	2184691	56.886 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022620\
Data File : FG006751.D
Signal(s) : FID1A.ch
Acq On : 26 Feb 2020 16:21
Operator : DD\AJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 27 01:22:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\C44-FG022520.M
Quant Title :
QLast Update : Tue Feb 25 16:35:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

