

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022819\
 Data File : FG003989.D
 Signal(s) : FID1A.ch
 Acq On : 28 Feb 2019 09:53
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 01 00:26:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
 Quant Title :
 QLast Update : Wed Feb 06 08:50:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.456	6781852	50.141 ug/ml
Target Compounds			
2) N-DECANE	4.182	8671474	54.978 ug/ml
3) N-DODECANE	6.206	7632061	48.339 ug/ml
4) N-TETRADECANE	8.001	8581822	54.373 ug/ml
5) N-HEXADECANE	9.601	8441283	53.584 ug/ml
6) N-OCTADECANE	11.045	8275201	52.568 ug/ml
7) N-EICOSANE	12.355	8010782	51.149 ug/ml
8) N-DOCOSANE	13.554	7757320	50.169 ug/ml
10) N-TETRACOSANE	14.659	7544420	50.071 ug/ml
11) N-HEXACOSANE	15.680	7458359	50.704 ug/ml
12) N-OCTACOSANE	16.633	7470189	51.850 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022819\
Data File : FG003989.D
Signal(s) : FID1A.ch
Acq On : 28 Feb 2019 09:53
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 01 00:26:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG020519.M
Quant Title :
QLast Update : Wed Feb 06 08:50:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

