

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022824\
 Data File : FG013561.D
 Signal(s) : FID1A.ch
 Acq On : 28 Feb 2024 12:35
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 28 12:50:04 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG022824.M
 Quant Title :
 QLast Update : Wed Feb 28 12:49:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.071	606311	5.123 ug/ml
Target Compounds			
1) N-OCTANE	2.057	643748	5.146 ug/ml
2) N-DECANE	4.584	652201	5.089 ug/ml
3) N-DODECANE	6.754	675376	5.062 ug/ml
4) N-TETRADECANE	8.588	661487	5.066 ug/ml
5) N-HEXADECANE	10.201	674819	5.079 ug/ml
6) N-OCTADECANE	11.649	695995	5.086 ug/ml
7) N-EICOSANE	12.964	679264	5.096 ug/ml
8) N-DOCOSANE	14.168	665771	5.095 ug/ml
10) N-TETRACOSANE	15.275	657974	5.099 ug/ml
11) N-HEXACOSANE	16.300	641951	5.098 ug/ml
12) N-OCTACOSANE	17.254	633674	5.127 ug/ml
13) N-TRIACONTANE	18.146	640526	5.236 ug/ml
14) N-DOTRIACONTANE	18.982	613554	5.257 ug/ml
15) N-TETRATRIACONTANE	19.769	491075	5.210 ug/ml
16) N-HEXATRIACONTANE	20.511	393474	5.268 ug/mlm
17) N-OCTATRIACONTANE	21.268	356146	5.165 ug/mlm
18) N-TETRACONTANE	22.217	392362	5.671 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG022824\
Data File : FG013561.D
Signal(s) : FID1A.ch
Acq On : 28 Feb 2024 12:35
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 28 12:50:04 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG022824.M
Quant Title :
QLast Update : Wed Feb 28 12:49:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

