

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG030519\
 Data File : FG004040.D
 Signal(s) : FID1A.ch
 Acq On : 05 Mar 2019 16:51
 Operator : AJ\SJ
 Sample : PB117621BSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB117621BSD

Manual Integrations
 APPROVED

Sohil
 3/6/2019 2:10:18 PM

Integration File: Sample.e
 Quant Time: Mar 06 00:37:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
 Quant Title :
 QLast Update : Tue Mar 05 03:05:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.441	2276013	18.104 ug/ml
Target Compounds			
1) N-OCTANE	2.096	1862553	14.171 ug/ml
2) N-DECANE	4.171	2286054	17.099 ug/ml
3) N-DODECANE	6.194	2345141	17.299 ug/ml
4) N-TETRADECANE	7.987	2422637	17.373 ug/ml
5) N-HEXADECANE	9.588	2464756	17.319 ug/ml
6) N-OCTADECANE	11.030	2480692	17.290 ug/ml
7) N-EICOSANE	12.340	2498777	17.492 ug/ml
8) N-DOCOSANE	13.539	2502176	17.628 ug/ml
10) N-TETRACOSANE	14.643	2528011	17.926 ug/ml
11) N-HEXACOSANE	15.665	2461798	17.676 ug/ml
12) N-OCTACOSANE	16.615	2413571	17.629 ug/ml
13) N-TRIACONTANE	17.504	2350458	17.600 ug/ml
14) N-DOTRIACONTANE	18.338	2241086	17.712 ug/ml
15) N-TETRATRIACONTANE	19.122	2034732	17.170 ug/ml
16) N-HEXATRIACONTANE	19.866	1878216	16.966 ug/mlm
17) N-OCTATRIACONTANE	20.580	1696664	16.808 ug/ml
18) N-TETRACONTANE	21.418	1845628	17.401 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG030519\
Data File : FG004040.D
Signal(s) : FID1A.ch
Acq On : 05 Mar 2019 16:51
Operator : AJ\SJ
Sample : PB117621BSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB117621BSD

Manual Integrations
APPROVED

Sohil
3/6/2019 2:10:18 PM

Integration File: Sample.e
Quant Time: Mar 06 00:37:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
Quant Title :
QLast Update : Tue Mar 05 03:05:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

