

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG030819\
 Data File : FG004105.D
 Signal(s) : FID1A.ch
 Acq On : 08 Mar 2019 17:38
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 08 22:34:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
 Quant Title :
 QLast Update : Tue Mar 05 03:05:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.441	5408621	43.021 ug/ml
Target Compounds			
2) N-DECANE	4.171	7245034	54.190 ug/ml
3) N-DODECANE	6.194	7232776	53.352 ug/ml
4) N-TETRADECANE	7.989	7153885	51.303 ug/ml
5) N-HEXADECANE	9.589	7021780	49.339 ug/ml
6) N-OCTADECANE	11.031	6837655	47.657 ug/ml
7) N-EICOSANE	12.342	6569449	45.987 ug/ml
8) N-DOCOSANE	13.540	6290535	44.318 ug/ml
10) N-TETRACOSANE	14.643	6062037	42.984 ug/ml
11) N-HEXACOSANE	15.665	5914341	42.466 ug/ml
12) N-OCTACOSANE	16.617	5888824	43.013 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG030819\
Data File : FG004105.D
Signal(s) : FID1A.ch
Acq On : 08 Mar 2019 17:38
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 08 22:34:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
Quant Title :
QLast Update : Tue Mar 05 03:05:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

