

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031424\
 Data File : FG013647.D
 Signal(s) : FID1A.ch
 Acq On : 14 Mar 2024 21:24
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 15 01:38:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG022824.M
 Quant Title :
 QLast Update : Wed Feb 28 13:13:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	15.071	6549307	55.336 ug/ml
Target Compounds				
2)	N-DECANE	4.584	7212625	56.277 ug/ml
3)	N-DODECANE	6.756	7451992	55.858 ug/ml
4)	N-TETRADECANE	8.589	7290395	55.832 ug/ml
5)	N-HEXADECANE	10.202	7401371	55.701 ug/ml
6)	N-OCTADECANE	11.651	7605869	55.583 ug/ml
7)	N-EICOSANE	12.966	7384269	55.395 ug/ml
8)	N-DOCOSANE	14.168	7230719	55.332 ug/ml
10)	N-TETRACOSANE	15.276	7137898	55.312 ug/ml
11)	N-HEXACOSANE	16.301	6971206	55.360 ug/ml
12)	N-OCTACOSANE	17.256	6893809	55.772 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031424\
Data File : FG013647.D
Signal(s) : FID1A.ch
Acq On : 14 Mar 2024 21:24
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 15 01:38:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG022824.M
Quant Title :
QLast Update : Wed Feb 28 13:13:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

