

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031521\
 Data File : FG007705.D
 Signal(s) : FID1A.ch
 Acq On : 15 Mar 2021 14:17
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 16 03:32:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031221.M
 Quant Title :
 QLast Update : Mon Mar 15 09:29:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.452	6262875	49.497 ug/ml
Target Compounds				
2)	N-DECANE	4.174	7451336	50.838 ug/ml
3)	N-DODECANE	6.193	7514522	50.733 ug/ml
4)	N-TETRADECANE	7.990	7652677	50.582 ug/ml
5)	N-HEXADECANE	9.593	7555621	50.281 ug/ml
6)	N-OCTADECANE	11.036	7544047	50.044 ug/ml
7)	N-EICOSANE	12.349	7378584	49.900 ug/ml
8)	N-DOCOSANE	13.550	7223187	49.635 ug/ml
10)	N-TETRACOSANE	14.657	7092849	49.436 ug/ml
11)	N-HEXACOSANE	15.680	6887921	49.286 ug/ml
12)	N-OCTACOSANE	16.632	6731941	49.226 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031521\
Data File : FG007705.D
Signal(s) : FID1A.ch
Acq On : 15 Mar 2021 14:17
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 16 03:32:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031221.M
Quant Title :
QLast Update : Mon Mar 15 09:29:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

