

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_G\Data\FG031619\  
 Data File : FG004229.D  
 Signal(s) : FID1A.ch  
 Acq On : 15 Mar 2019 21:37  
 Operator : AJ\SJ  
 Sample : 50 PPM TRPH STD  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 FID\_G  
 ClientSampleId :  
 50 PPM TRPH STD

Integration File: autoint1.e  
 Quant Time: Mar 15 23:58:10 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_G\Method\FG030419.M  
 Quant Title :  
 QLast Update : Tue Mar 05 03:05:54 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1uL  
 Signal Phase : Rxi-1ms  
 Signal Info : 20mx0.18mmx0.18um

| Compound                      | R.T.   | Response | Conc Units   |
|-------------------------------|--------|----------|--------------|
| -----                         |        |          |              |
| System Monitoring Compounds   |        |          |              |
| 9) S TETRACOSANE-d50 (SURR... | 14.446 | 5329965  | 42.396 ug/ml |
| Target Compounds              |        |          |              |
| 2) N-DECANE                   | 4.168  | 6928033  | 51.818 ug/ml |
| 3) N-DODECANE                 | 6.193  | 6878216  | 50.736 ug/ml |
| 4) N-TETRADECANE              | 7.989  | 6789657  | 48.691 ug/ml |
| 5) N-HEXADECANE               | 9.591  | 6653220  | 46.749 ug/ml |
| 6) N-OCTADECANE               | 11.033 | 6491434  | 45.244 ug/ml |
| 7) N-EICOSANE                 | 12.345 | 6282592  | 43.979 ug/ml |
| 8) N-DOCOSANE                 | 13.546 | 6154699  | 43.361 ug/ml |
| 10) N-TETRACOSANE             | 14.651 | 6101870  | 43.267 ug/ml |
| 11) N-HEXACOSANE              | 15.672 | 6083895  | 43.683 ug/ml |
| 12) N-OCTACOSANE              | 16.625 | 6181083  | 45.147 ug/ml |
| -----                         |        |          |              |

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_G\Data\FG031619\  
Data File : FG004229.D  
Signal(s) : FID1A.ch  
Acq On : 15 Mar 2019 21:37  
Operator : AJ\SJ  
Sample : 50 PPM TRPH STD  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
FID\_G  
ClientSampleId :  
50 PPM TRPH STD

Integration File: autoint1.e  
Quant Time: Mar 15 23:58:10 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_G\Method\FG030419.M  
Quant Title :  
QLast Update : Tue Mar 05 03:05:54 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um

