

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031921\
 Data File : FG007730.D
 Signal(s) : FID1A.ch
 Acq On : 19 Mar 2021 15:01
 Operator : DD\AJ
 Sample : M1707-05MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 19 22:47:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031221.M
 Quant Title :
 QLast Update : Mon Mar 15 09:29:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.447	2504180	19.791 ug/ml
Target Compounds			
1) N-OCTANE	2.106	2964164	20.635 ug/ml
2) N-DECANE	4.172	3460835	23.612 ug/ml
3) N-DODECANE	6.191	3493380	23.585 ug/ml
4) N-TETRADECANE	7.986	3599782	23.793 ug/ml
5) N-HEXADECANE	9.589	3488623	23.216 ug/ml
6) N-OCTADECANE	11.033	3465005	22.986 ug/ml
7) N-EICOSANE	12.344	3461721	23.411 ug/ml
8) N-DOCOSANE	13.546	3374453	23.188 ug/ml
10) N-TETRACOSANE	14.652	3280967	22.868 ug/ml
11) N-HEXACOSANE	15.674	3187244	22.806 ug/ml
12) N-OCTACOSANE	16.625	3035012	22.193 ug/ml
13) N-TRIACONTANE	17.517	2957829	22.363 ug/ml
14) N-DOTRIACONTANE	18.350	2842333	22.562 ug/ml
15) N-TETRATRIACONTANE	19.137	2517340	21.933 ug/ml
16) N-HEXATRIACONTANE	19.880	2333501	23.062 ug/ml
17) N-OCTATRIACONTANE	20.592	2276117	24.855 ug/ml
18) N-TETRACONTANE	21.431	2256986	25.772 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG031921\
Data File : FG007730.D
Signal(s) : FID1A.ch
Acq On : 19 Mar 2021 15:01
Operator : DD\AJ
Sample : M1707-05MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 19 22:47:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031221.M
Quant Title :
QLast Update : Mon Mar 15 09:29:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

