

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032023\
 Data File : FG011619.D
 Signal(s) : FID1A.ch
 Acq On : 20 Mar 2023 09:59
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 21 04:15:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
 Quant Title :
 QLast Update : Wed Mar 15 18:43:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.054	5095490	50.624 ug/ml
Target Compounds			
1) N-OCTANE	2.064	5429914	46.138 ug/ml
2) N-DECANE	4.581	5792235	47.927 ug/ml
3) N-DODECANE	6.751	5902081	48.368 ug/ml
4) N-TETRADECANE	8.582	5939391	48.808 ug/ml
5) N-HEXADECANE	10.193	6050949	49.071 ug/ml
6) N-OCTADECANE	11.639	6060589	49.621 ug/ml
7) N-EICOSANE	12.952	6025212	50.144 ug/ml
8) N-DOCOSANE	14.154	5803341	50.591 ug/ml
10) N-TETRACOSANE	15.259	5791304	51.269 ug/ml
11) N-HEXACOSANE	16.282	5547964	51.374 ug/ml
12) N-OCTACOSANE	17.234	5337388	51.473 ug/ml
13) N-TRIACONTANE	18.123	5114061	51.605 ug/ml
14) N-DOTRIACONTANE	18.957	4826387	52.039 ug/ml
15) N-TETRATRIACONTANE	19.743	4553880	54.016 ug/ml
16) N-HEXATRIACONTANE	20.486	4558808	59.217 ug/ml
17) N-OCTATRIACONTANE	21.239	4817144	63.982 ug/ml
18) N-TETRACONTANE	22.180	4987246	67.952 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032023\
Data File : FG011619.D
Signal(s) : FID1A.ch
Acq On : 20 Mar 2023 09:59
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 21 04:15:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
Quant Title :
QLast Update : Wed Mar 15 18:43:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

