

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032023\
 Data File : FG011632.D
 Signal(s) : FID1A.ch
 Acq On : 20 Mar 2023 18:07
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 21 04:17:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
 Quant Title :
 QLast Update : Wed Mar 15 18:43:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.055	4640551	46.104 ug/ml
Target Compounds			
1) N-OCTANE	2.065	5031230	42.750 ug/ml
2) N-DECANE	4.581	5373002	44.459 ug/ml
3) N-DODECANE	6.752	5468777	44.817 ug/ml
4) N-TETRADECANE	8.582	5498833	45.188 ug/ml
5) N-HEXADECANE	10.194	5611744	45.509 ug/ml
6) N-OCTADECANE	11.640	5613897	45.964 ug/ml
7) N-EICOSANE	12.954	5549400	46.184 ug/ml
8) N-DOCOSANE	14.154	5317293	46.354 ug/ml
10) N-TETRACOSANE	15.260	5270884	46.661 ug/ml
11) N-HEXACOSANE	16.284	5045135	46.718 ug/ml
12) N-OCTACOSANE	17.236	4900955	47.264 ug/ml
13) N-TRIACONTANE	18.125	4749464	47.926 ug/ml
14) N-DOTRIACONTANE	18.960	4526203	48.803 ug/ml
15) N-TETRATRIACONTANE	19.745	4251071	50.424 ug/ml
16) N-HEXATRIACONTANE	20.487	4202034	54.583 ug/ml
17) N-OCTATRIACONTANE	21.238	4459546	59.232 ug/ml
18) N-TETRACONTANE	22.178	4701454	64.058 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032023\
Data File : FG011632.D
Signal(s) : FID1A.ch
Acq On : 20 Mar 2023 18:07
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 21 04:17:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
Quant Title :
QLast Update : Wed Mar 15 18:43:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

