

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032321\
 Data File : FG007751.D
 Signal(s) : FID1A.ch
 Acq On : 23 Mar 2021 16:36
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 24 01:40:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031221.M
 Quant Title :
 QLast Update : Mon Mar 15 09:29:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.455	5646072	44.622 ug/ml
Target Compounds			
1) N-OCTANE	2.101	6822809	47.496 ug/ml
2) N-DECANE	4.175	7275858	49.641 ug/ml
3) N-DODECANE	6.195	7239229	48.874 ug/ml
4) N-TETRADECANE	7.993	7235383	47.824 ug/ml
5) N-HEXADECANE	9.595	7028034	46.770 ug/ml
6) N-OCTADECANE	11.039	6927611	45.955 ug/ml
7) N-EICOSANE	12.351	6700435	45.314 ug/ml
8) N-DOCOSANE	13.553	6491760	44.609 ug/ml
10) N-TETRACOSANE	14.659	6359371	44.323 ug/ml
11) N-HEXACOSANE	15.682	6203816	44.391 ug/ml
12) N-OCTACOSANE	16.634	6191203	45.272 ug/ml
13) N-TRIACONTANE	17.527	6208383	46.940 ug/ml
14) N-DOTRIACONTANE	18.361	6150540	48.822 ug/ml
15) N-TETRATRIACONTANE	19.146	5959465	51.923 ug/ml
16) N-HEXATRIACONTANE	19.889	5571029	55.058 ug/mlm
17) N-OCTATRIACONTANE	20.602	5419943	59.185 ug/ml
18) N-TETRACONTANE	21.445	5200337	59.380 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032321\
Data File : FG007751.D
Signal(s) : FID1A.ch
Acq On : 23 Mar 2021 16:36
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 24 01:40:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031221.M
Quant Title :
QLast Update : Mon Mar 15 09:29:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

