

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/27/2023
Supervised By :Ankita Jodhani 03/27/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG03242
Data File : FG011659.D
Signal(s) : FID1A.ch
Acq On : 24 Mar 2023 10:56
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.057	2.013	2.154	BB	395518	5156224	87.71%	5.439%
2	4.576	4.543	4.631	BB	532057	5563703	94.64%	5.868%
3	6.747	6.707	6.797	BB	554914	5654832	96.19%	5.965%
4	8.577	8.543	8.631	BB	547655	5695367	96.88%	6.007%
5	10.188	10.148	10.243	BB	537248	5842338	99.38%	6.162%
6	11.634	11.598	11.680	BB	504492	5879018	100.00%	6.201%
7	12.947	12.894	12.998	BB	485387	5850405	99.51%	6.171%
8	14.148	14.114	14.198	VB	458015	5648633	96.08%	5.958%
9	15.049	14.992	15.074	BV	361377	4958915	84.35%	5.230%
10	15.254	15.185	15.291	VB	442931	5678248	96.58%	5.989%
11	16.277	16.226	16.329	BB	407062	5449148	92.69%	5.748%
12	17.229	17.154	17.293	BB	406732	5314913	90.40%	5.606%
13	18.118	18.064	18.158	BB	373941	5142940	87.48%	5.425%
14	18.953	18.893	19.013	BB	339635	4874142	82.91%	5.141%
15	19.738	19.688	19.776	BB	321336	4500465	76.55%	4.747%
16	20.479	20.435	20.523	M	310441	4344548	73.90%	4.582%
17	21.228	21.172	21.293	BB	240995	4527860	77.02%	4.776%
18	22.166	22.076	22.261	BB	196370	4726432	80.39%	4.985%
Sum of corrected areas:							94808129	

FG031523.M Fri Mar 24 18:20:13 2023