

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032519\
 Data File : FG004343.D
 Signal(s) : FID1A.ch
 Acq On : 25 Mar 2019 10:27
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 25 14:46:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
 Quant Title :
 QLast Update : Tue Mar 05 03:05:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.432	6201868	49.331 ug/ml
Target Compounds			
1) N-OCTANE	2.078	6998098	53.246 ug/ml
2) N-DECANE	4.161	7777902	58.175 ug/ml
3) N-DODECANE	6.185	7772224	57.331 ug/ml
4) N-TETRADECANE	7.979	7713987	55.319 ug/ml
5) N-HEXADECANE	9.581	7604854	53.436 ug/ml
6) N-OCTADECANE	11.023	7461638	52.007 ug/ml
7) N-EICOSANE	12.333	7246486	50.726 ug/ml
8) N-DOCOSANE	13.532	7078042	49.866 ug/ml
10) N-TETRACOSANE	14.637	7010533	49.710 ug/ml
11) N-HEXACOSANE	15.659	6966129	50.018 ug/ml
12) N-OCTACOSANE	16.610	7127096	52.057 ug/ml
13) N-TRIACONTANE	17.498	7398034	55.397 ug/ml
14) N-DOTRIACONTANE	18.335	7377807	58.308 ug/ml
15) N-TETRATRIACONTANE	19.121	7336603	61.911 ug/mlm
16) N-HEXATRIACONTANE	19.861	7084200	63.991 ug/mlm
17) N-OCTATRIACONTANE	20.575	6533248	64.720 ug/ml
18) N-TETRACONTANE	21.409	6930175	65.340 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032519\
Data File : FG004343.D
Signal(s) : FID1A.ch
Acq On : 25 Mar 2019 10:27
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 25 14:46:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG030419.M
Quant Title :
QLast Update : Tue Mar 05 03:05:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

