

rteres

Instrument :

FID_G

LabSampleId :

100 TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/26/2022

Supervised By :mohammad ahmed 03/30/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032522
Data File : FG009401.D
Signal(s) : FID1A.ch
Acq On : 25 Mar 2022 19:49
Sample : 100 TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.233	2.205	2.286	BV	1732957	13004666	90.42%	5.828%
2	4.331	4.296	4.400	BB	1386176	13303430	92.50%	5.962%
3	6.358	6.308	6.422	BB	1266398	13536142	94.12%	6.066%
4	8.158	8.104	8.218	BB	1253463	13815739	96.06%	6.192%
5	9.762	9.708	9.821	BB	1195020	13976345	97.18%	6.264%
6	11.207	11.168	11.269	VB	1141657	14381999	100.00%	6.445%
7	12.521	12.468	12.573	BB	1079983	14017601	97.47%	6.282%
8	13.723	13.637	13.778	BB	1062428	13907310	96.70%	6.233%
9	14.624	14.573	14.692	BV	842858	12369303	86.01%	5.543%
10	14.831	14.765	14.884	BB	1022339	13837478	96.21%	6.201%
11	15.853	15.790	15.899	BB	967877	13614181	94.66%	6.101%
12	16.806	16.715	16.868	BB	926964	13297694	92.46%	5.959%
13	17.696	17.628	17.753	BB	863346	12740054	88.58%	5.710%
14	18.530	18.458	18.595	BB	812623	11892409	82.69%	5.330%
15	19.314	19.249	19.361	BB	684086	10560466	73.43%	4.733%
16	20.054	20.000	20.090	BBA	617337	8963781	62.33%	4.017%
17	20.790	20.718	20.865	BB	446661	7986895	55.53%	3.579%
18	21.695	21.606	21.841	BV	336440	7930220	55.14%	3.554%
Sum of corrected areas:						223135710		

FG032522.M Sat Mar 26 11:16:54 2022