

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032622\
 Data File : FG009409.D
 Signal(s) : FID1A.ch
 Acq On : 26 Mar 2022 07:33
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/28/2022
 Supervised By :mohammad ahmed 03/30/2022

Integration File: autoint1.e
 Quant Time: Mar 26 22:45:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
 Quant Title :
 QLast Update : Sat Mar 26 00:33:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.621	6520867	47.319 ug/ml
Target Compounds			
1) N-OCTANE	2.232	6827147	47.471 ug/ml
2) N-DECANE	4.329	7024360	48.021 ug/ml
3) N-DODECANE	6.355	7135517	47.848 ug/ml
4) N-TETRADECANE	8.153	7272241	47.631 ug/ml
5) N-HEXADECANE	9.759	7343697	47.403 ug/ml
6) N-OCTADECANE	11.205	7549554	47.311 ug/ml
7) N-EICOSANE	12.518	7369425	47.204 ug/ml
8) N-DOCOSANE	13.719	7323271	47.313 ug/ml
10) N-TETRACOSANE	14.826	7296020	47.309 ug/ml
11) N-HEXACOSANE	15.849	7187624	47.343 ug/ml
12) N-OCTACOSANE	16.801	7029436	47.432 ug/ml
13) N-TRIACONTANE	17.690	6748536	47.323 ug/ml
14) N-DOTRIACONTANE	18.523	6294897	47.339 ug/ml
15) N-TETRATRIACONTANE	19.309	5555821	47.549 ug/ml
16) N-HEXATRIACONTANE	20.050	4669394	47.349 ug/ml
17) N-OCTATRIACONTANE	20.788	4120852	47.536 ug/ml
18) N-TETRACONTANE	21.694	4054817	47.704 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032622\
Data File : FG009409.D
Signal(s) : FID1A.ch
Acq On : 26 Mar 2022 07:33
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2022
Supervised By :mohammad ahmed 03/30/2022

Integration File: autoint1.e
Quant Time: Mar 26 22:45:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
Quant Title :
QLast Update : Sat Mar 26 00:33:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

