

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032622\
 Data File : FG009436.D
 Signal(s) : FID1A.ch
 Acq On : 26 Mar 2022 22:07
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/28/2022
 Supervised By :mohammad ahmed 03/30/2022

Integration File: autoint1.e
 Quant Time: Mar 26 22:48:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
 Quant Title :
 QLast Update : Sat Mar 26 00:33:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.632	5606356	40.683 ug/ml
Target Compounds			
1) N-OCTANE	2.234	6989706	48.601 ug/ml
2) N-DECANE	4.333	7154129	48.908 ug/ml
3) N-DODECANE	6.360	7123068	47.764 ug/ml
4) N-TETRADECANE	8.160	7086008	46.411 ug/ml
5) N-HEXADECANE	9.767	6976621	45.034 ug/ml
6) N-OCTADECANE	11.214	6977989	43.729 ug/ml
7) N-EICOSANE	12.527	6631619	42.478 ug/ml
8) N-DOCOSANE	13.728	6413500	41.435 ug/ml
10) N-TETRACOSANE	14.836	6269552	40.653 ug/ml
11) N-HEXACOSANE	15.861	6084006	40.074 ug/ml
12) N-OCTACOSANE	16.814	5963774	40.242 ug/ml
13) N-TRIACONTANE	17.705	5838313	40.940 ug/ml
14) N-DOTRIACONTANE	18.540	5663213	42.589 ug/ml
15) N-TETRATRIACONTANE	19.327	5277564	45.168 ug/mlm
16) N-HEXATRIACONTANE	20.068	4695808	47.617 ug/mlm
17) N-OCTATRIACONTANE	20.807	4349177	50.169 ug/ml
18) N-TETRACONTANE	21.724	4258272	50.098 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032622\
Data File : FG009436.D
Signal(s) : FID1A.ch
Acq On : 26 Mar 2022 22:07
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2022
Supervised By :mohammad ahmed 03/30/2022

Integration File: autoint1.e
Quant Time: Mar 26 22:48:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
Quant Title :
QLast Update : Sat Mar 26 00:33:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

