

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032923\
 Data File : FG011685.D
 Signal(s) : FID1A.ch
 Acq On : 29 Mar 2023 13:12
 Operator : YP\AJ
 Sample : PB151763BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 PB151763BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/30/2023
 Supervised By :Ankita Jodhani 03/30/2023

Integration File: autoint1.e
 Quant Time: Mar 29 18:00:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
 Quant Title :
 QLast Update : Wed Mar 15 18:43:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.046	2137572	21.237 ug/ml
Target Compounds			
1) N-OCTANE	2.056	2408564	20.466 ug/ml
2) N-DECANE	4.574	2445038	20.231 ug/ml
3) N-DODECANE	6.744	2426762	19.887 ug/ml
4) N-TETRADECANE	8.574	2382979	19.583 ug/ml
5) N-HEXADECANE	10.185	2432908	19.730 ug/ml
6) N-OCTADECANE	11.632	2448438	20.047 ug/ml
7) N-EICOSANE	12.944	2447364	20.368 ug/ml
8) N-DOCOSANE	14.144	2372418	20.682 ug/ml
10) N-TETRACOSANE	15.251	2392373	21.179 ug/ml
11) N-HEXACOSANE	16.274	2324322	21.523 ug/ml
12) N-OCTACOSANE	17.226	2262423	21.818 ug/ml
13) N-TRIACONTANE	18.116	2211491	22.316 ug/ml
14) N-DOTRIACONTANE	18.949	2119438	22.852 ug/ml
15) N-TETRATRIACONTANE	19.735	2024286	24.011 ug/ml
16) N-HEXATRIACONTANE	20.477	2038072	26.474 ug/mlm
17) N-OCTATRIACONTANE	21.225	2131826	28.315 ug/ml
18) N-TETRACONTANE	22.159	2193273	29.884 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032923\
Data File : FG011685.D
Signal(s) : FID1A.ch
Acq On : 29 Mar 2023 13:12
Operator : YP\AJ
Sample : PB151763BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
PB151763BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/30/2023
Supervised By :Ankita Jodhani 03/30/2023

Integration File: autoint1.e
Quant Time: Mar 29 18:00:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031523.M
Quant Title :
QLast Update : Wed Mar 15 18:43:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

