

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032924\
 Data File : FG013668.D
 Signal(s) : FID1A.ch
 Acq On : 29 Mar 2024 14:14
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 30 01:00:28 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031524.M
 Quant Title :
 QLast Update : Fri Mar 15 13:14:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.074	2083730	17.913 ug/ml
Target Compounds			
15) N-TETRATRIACONTANE	19.775	61810	0.656 ug/ml
18) N-TETRACONTANE	22.143	66136	1.044 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032924\
Data File : FG013668.D
Signal(s) : FID1A.ch
Acq On : 29 Mar 2024 14:14
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 30 01:00:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031524.M
Quant Title :
QLast Update : Fri Mar 15 13:14:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 15.074 min
Delta R.T.: 0.007 min
Response: 2083730
Conc: 17.91 ug/ml

Instrument :


FID_G

ClientSampleId :



#15 N-TETRATRIACONTANE

R.T.: 19.775 min
Delta R.T.: 0.013 min
Response: 61810
Conc: 0.66 ug/ml



#18 N-TETRACONTANE

R.T.: 22.143 min
Delta R.T.: -0.056 min
Response: 66136
Conc: 1.04 ug/ml