

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032924\
 Data File : FG013676.D
 Signal(s) : FID1A.ch
 Acq On : 29 Mar 2024 18:13
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 30 01:01:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031524.M
 Quant Title :
 QLast Update : Fri Mar 15 13:14:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.073	5500783	47.288 ug/ml
Target Compounds			
2) N-DECANE	4.584	5873781	45.620 ug/ml
3) N-DODECANE	6.755	6106497	45.704 ug/ml
4) N-TETRADECANE	8.589	6009586	46.216 ug/ml
5) N-HEXADECANE	10.203	6148479	46.690 ug/ml
6) N-OCTADECANE	11.652	6351574	46.984 ug/ml
7) N-EICOSANE	12.967	6197016	47.215 ug/ml
8) N-DOCOSANE	14.171	6098137	47.393 ug/ml
10) N-TETRACOSANE	15.279	6032142	47.412 ug/ml
11) N-HEXACOSANE	16.305	5861871	47.295 ug/ml
12) N-OCTACOSANE	17.258	5782826	47.374 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG032924\
Data File : FG013676.D
Signal(s) : FID1A.ch
Acq On : 29 Mar 2024 18:13
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 30 01:01:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG031524.M
Quant Title :
QLast Update : Fri Mar 15 13:14:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

