

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG033022\
 Data File : FG009474.D
 Signal(s) : FID1A.ch
 Acq On : 30 Mar 2022 12:04
 Operator : AJ\MA
 Sample : N1645-12
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 MDL-WATER-06-QT1-2022

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/31/2022
 Supervised By :mohammad ahmed 04/01/2022

Integration File: autoint1.e
 Quant Time: Mar 31 02:00:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
 Quant Title :
 QLast Update : Sat Mar 26 00:33:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.628	2224836	16.145 ug/mlm
Target Compounds			
1) N-OCTANE	2.234	139175	0.968 ug/ml
2) N-DECANE	4.329	143964	0.984 ug/ml
3) N-DODECANE	6.355	142409	0.955 ug/ml
4) N-TETRADECANE	8.154	145895	0.956 ug/ml
5) N-HEXADECANE	9.760	148746	0.960 ug/ml
6) N-OCTADECANE	11.207	152045	0.953 ug/ml
7) N-EICOSANE	12.521	150337	0.963 ug/ml
8) N-DOCOSANE	13.724	146706	0.948 ug/ml
10) N-TETRACOSANE	14.831	144077	0.934 ug/ml
11) N-HEXACOSANE	15.855	138531	0.912 ug/ml
12) N-OCTACOSANE	16.810	141673	0.956 ug/ml
13) N-TRIACONTANE	17.700	155808	1.093 ug/ml
14) N-DOTRIACONTANE	18.536	154393	1.161 ug/ml
15) N-TETRATRIACONTANE	19.322	136841	1.171 ug/ml
16) N-HEXATRIACONTANE	20.065	125278	1.270 ug/mlm
17) N-OCTATRIACONTANE	20.804	98621	1.138 ug/ml
18) N-TETRACONTANE	21.719	101788	1.198 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG033022\
Data File : FG009474.D
Signal(s) : FID1A.ch
Acq On : 30 Mar 2022 12:04
Operator : AJ\MA
Sample : N1645-12
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
MDL-WATER-06-QT1-2022

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/31/2022
Supervised By :mohammad ahmed 04/01/2022

Integration File: autoint1.e
Quant Time: Mar 31 02:00:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032522.M
Quant Title :
QLast Update : Sat Mar 26 00:33:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

