

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040319\
 Data File : FG004556.D
 Signal(s) : FID1A.ch
 Acq On : 03 Apr 2019 15:33
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Apr 04 03:18:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032719.M
 Quant Title :
 QLast Update : Wed Mar 27 13:30:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.420	4865147	47.167 ug/ml
Target Compounds			
2) N-DECANE	4.151	6287620	59.567 ug/ml
3) N-DODECANE	6.174	6306965	59.862 ug/ml
4) N-TETRADECANE	7.968	6259103	57.769 ug/ml
5) N-HEXADECANE	9.569	6146669	54.939 ug/ml
6) N-OCTADECANE	11.011	5981896	52.739 ug/ml
7) N-EICOSANE	12.321	5795357	50.696 ug/ml
8) N-DOCOSANE	13.522	5636010	49.304 ug/ml
10) N-TETRACOSANE	14.625	5570837	48.226 ug/ml
11) N-HEXACOSANE	15.648	5493843	47.925 ug/ml
12) N-OCTACOSANE	16.598	5542809	47.912 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040319\
Data File : FG004556.D
Signal(s) : FID1A.ch
Acq On : 03 Apr 2019 15:33
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Apr 04 03:18:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG032719.M
Quant Title :
QLast Update : Wed Mar 27 13:30:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

