

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040521\
 Data File : FG007775.D
 Signal(s) : FID1A.ch
 Acq On : 06 Apr 2021 00:04
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Ankita
 4/6/2021 2:57:04 PM

Integration File: autoint1.e
 Quant Time: Apr 06 02:58:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
 Quant Title :
 QLast Update : Tue Apr 06 02:57:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.772	748271	5.150 ug/ml
Target Compounds			
1) N-OCTANE	2.376	816298	5.319 ug/ml
2) N-DECANE	4.483	844307	5.287 ug/ml
3) N-DODECANE	6.510	853055	5.271 ug/ml
4) N-TETRADECANE	8.310	873070	5.256 ug/ml
5) N-HEXADECANE	9.914	872211	5.260 ug/ml
6) N-OCTADECANE	11.358	888331	5.277 ug/ml
7) N-EICOSANE	12.670	878390	5.249 ug/ml
8) N-DOCOSANE	13.871	870117	5.240 ug/ml
10) N-TETRACOSANE	14.976	860757	5.203 ug/ml
11) N-HEXACOSANE	15.997	832095	5.142 ug/ml
12) N-OCTACOSANE	16.949	810664	5.114 ug/ml
13) N-TRIACONTANE	17.835	795364	5.217 ug/ml
14) N-DOTRIACONTANE	18.669	821338	5.733 ug/mlm
15) N-TETRATRIACONTANE	19.450	661006	5.067 ug/ml
16) N-HEXATRIACONTANE	20.190	571800	4.995 ug/ml
17) N-OCTATRIACONTANE	20.949	531504	5.180 ug/ml
18) N-TETRACONTANE	21.907	498230	5.220 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040521\
Data File : FG007775.D
Signal(s) : FID1A.ch
Acq On : 06 Apr 2021 00:04
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_G
ClientSampled :
5 TRPH STD

Manual Integrations
APPROVED

Ankita
4/6/2021 2:57:04 PM

Integration File: autoint1.e
Quant Time: Apr 06 02:58:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
Quant Title :
QLast Update : Tue Apr 06 02:57:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

