

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040721\
 Data File : FG007793.D
 Signal(s) : FID1A.ch
 Acq On : 07 Apr 2021 10:27
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 08 02:06:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
 Quant Title :
 QLast Update : Tue Apr 06 03:05:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.777	7851763	52.578 ug/ml
Target Compounds			
1) N-OCTANE	2.370	8108640	50.662 ug/ml
2) N-DECANE	4.483	8675783	52.327 ug/ml
3) N-DODECANE	6.511	8818112	52.553 ug/ml
4) N-TETRADECANE	8.312	9061919	52.657 ug/ml
5) N-HEXADECANE	9.916	9052764	52.676 ug/ml
6) N-OCTADECANE	11.362	9179359	52.576 ug/ml
7) N-EICOSANE	12.674	9096006	52.474 ug/ml
8) N-DOCOSANE	13.876	8996147	52.266 ug/ml
10) N-TETRACOSANE	14.981	8888928	52.007 ug/ml
11) N-HEXACOSANE	16.002	8636667	51.892 ug/ml
12) N-OCTACOSANE	16.953	8412778	51.724 ug/ml
13) N-TRIACONTANE	17.841	8063363	51.316 ug/ml
14) N-DOTRIACONTANE	18.672	7550048	49.459 ug/ml
15) N-TETRATRIACONTANE	19.456	6871519	51.690 ug/ml
16) N-HEXATRIACONTANE	20.197	6040992	51.904 ug/ml
17) N-OCTATRIACONTANE	20.954	5436513	51.906 ug/ml
18) N-TETRACONTANE	21.910	5191783	52.370 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040721\
Data File : FG007793.D
Signal(s) : FID1A.ch
Acq On : 07 Apr 2021 10:27
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 08 02:06:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040521.M
Quant Title :
QLast Update : Tue Apr 06 03:05:53 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

