

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040722\
 Data File : FG009549.D
 Signal(s) : FID1A.ch
 Acq On : 07 Apr 2022 21:25
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/08/2022
 Supervised By :mohammad ahmed 04/12/2022

Integration File: Sample.e
 Quant Time: Apr 08 02:30:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
 Quant Title :
 QLast Update : Fri Apr 01 13:54:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.648	5711658	44.336 ug/ml
Target Compounds			
1) N-OCTANE	2.233	7266974	49.740 ug/ml
2) N-DECANE	4.333	7330718	49.374 ug/ml
3) N-DODECANE	6.363	7286518	48.347 ug/ml
4) N-TETRADECANE	8.166	7233180	47.220 ug/ml
5) N-HEXADECANE	9.774	7119942	46.138 ug/ml
6) N-OCTADECANE	11.223	7113768	45.214 ug/ml
7) N-EICOSANE	12.539	6763848	44.659 ug/ml
8) N-DOCOSANE	13.743	6532437	44.265 ug/ml
10) N-TETRACOSANE	14.853	6369828	44.434 ug/ml
11) N-HEXACOSANE	15.880	6207993	45.192 ug/mlm
12) N-OCTACOSANE	16.834	6082790	46.608 ug/mlm
13) N-TRIACONTANE	17.727	6017713	48.734 ug/ml
14) N-DOTRIACONTANE	18.564	5892170	50.760 ug/mlm
15) N-TETRATRIACONTANE	19.354	5576371	52.444 ug/mlm
16) N-HEXATRIACONTANE	20.096	4906440	51.814 ug/mlm
17) N-OCTATRIACONTANE	20.840	4429243	51.436 ug/mlm
18) N-TETRACONTANE	21.767	4564519	52.888 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG040722\
Data File : FG009549.D
Signal(s) : FID1A.ch
Acq On : 07 Apr 2022 21:25
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/12/2022

Integration File: Sample.e
Quant Time: Apr 08 02:30:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
Quant Title :
QLast Update : Fri Apr 01 13:54:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

