

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/08/2022
Supervised By :mohammad ahmed 04/12/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG04072
Data File : FG009538.D
Signal(s) : FID1A.ch
Acq On : 07 Apr 2022 12:38
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.231	2.196	2.287	BB	1080251	8107661	92.15%	6.202%
2	4.327	4.293	4.415	BB	874596	8663238	98.46%	6.626%
3	6.356	6.317	6.433	BB	795764	8734981	99.28%	6.681%
4	8.156	8.118	8.234	BB	753683	8798405	100.00%	6.730%
5	9.764	9.717	9.821	BB	719516	8729687	99.22%	6.677%
6	11.212	11.174	11.280	VB	715157	8771056	99.69%	6.709%
7	12.527	12.483	12.581	BB	650027	8295530	94.28%	6.345%
8	13.730	13.692	13.786	VB	623261	7944035	90.29%	6.076%
9	14.633	14.582	14.664	BV	476970	6835524	77.69%	5.228%
10	14.838	14.758	14.898	BB	560482	7603000	86.41%	5.815%
11	15.864	15.804	15.922	BB	521851	7190928	81.73%	5.500%
12	16.817	16.744	16.872	BB	497172	6889490	78.30%	5.270%
13	17.708	17.648	17.749	BB	461210	6690316	76.04%	5.117%
14	18.546	18.483	18.595	BB	438393	6461666	73.44%	4.942%
15	19.332	19.262	19.394	BV	397503	6013614	68.35%	4.600%
16	20.073	20.025	20.116	M	342150	5311414	60.37%	4.063%
17	20.815	20.674	20.895	BB	264604	4872902	55.38%	3.727%
18	21.732	21.639	21.815	BB	190406	4823609	54.82%	3.690%
Sum of corrected areas:						130737053		

FG040122.M Fri Apr 08 11:10:37 2022