

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/08/2022

Supervised By :mohammad ahmed 04/12/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG04072
Data File : FG009549.D
Signal(s) : FID1A.ch
Acq On : 07 Apr 2022 21:25
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG040122.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.233	2.198	2.291	BB	971507	7266974	99.13%	6.463%
2	4.333	4.294	4.423	BB	733097	7330718	100.00%	6.520%
3	6.363	6.330	6.453	BB	651410	7286518	99.40%	6.481%
4	8.166	8.121	8.250	BB	623522	7233180	98.67%	6.433%
5	9.774	9.727	9.846	BB	593922	7119942	97.12%	6.333%
6	11.223	11.186	11.286	BB	581267	7113768	97.04%	6.327%
7	12.539	12.447	12.605	BB	537962	6763848	92.27%	6.016%
8	13.743	13.667	13.801	BB	483698	6532437	89.11%	5.810%
9	14.648	14.558	14.718	BB	397629	5711658	77.91%	5.080%
10	14.853	14.775	14.908	BB	465211	6369828	86.89%	5.665%
11	15.879	15.719	15.923	BB	447302	6213978	84.77%	5.527%
12	16.834	16.762	16.885	M	441248	6139917	83.76%	5.461%
13	17.727	17.640	17.763	PV	418098	6017713	82.09%	5.352%
14	18.564	18.516	18.612	M	403261	5843627	79.71%	5.197%
15	19.354	19.295	19.384	M	341819	5395611	73.60%	4.799%
16	20.096	20.050	20.128	M	310466	4832818	65.93%	4.298%
17	20.840	20.795	20.889	M	233878	4566637	62.29%	4.062%
18	21.767	21.707	21.820	M	175800	4695327	64.05%	4.176%
Sum of corrected areas:						112434500		

FG040122.M Fri Apr 08 11:13:58 2022